



ELISA Technologies, Inc.

CERTIFICATE OF ANALYSIS

KIT TYPE: EZ Gluten
LOT NUMBER: EZG 25020-102

CATALOG NUMBER: 510EZG
EXPIRATION DATE: FEB/2026

KIT COMPONENTS

| COMPONENT | LOT NUMBER | EXPIRATION DATE |
|---------------------|------------|-----------------|
| TEST STRIP | EZG25020S | FEB/2026 |
| EXTRACTION SOLUTION | ES25012 | FEB/2026 |

CONTROL DATA

| SPECIFICATION | | VISUAL RESULT | PASS/FAIL |
|-----------------------|---------|---------------|-----------|
| NEGATIVE CONTROL | 0 ppm | NEGATIVE | PASS |
| LOW POSITIVE CONTROL | 5 ppm | POSITIVE | PASS |
| POSITIVE CONTROL | 10 ppm | POSITIVE | PASS |
| HIGH POSITIVE CONTROL | 500 ppm | POSITIVE | PASS |
| BROWN RICE FLOUR | | NEGATIVE | PASS |
| OAT BRAN | | NEGATIVE | PASS |
| COOKED EGG | | NEGATIVE | PASS |
| PEANUT | | NEGATIVE | PASS |
| CORN MEAL | | NEGATIVE | PASS |
| POWDERED MILK | | NEGATIVE | PASS |
| SOY FLOUR | | NEGATIVE | PASS |
| SUGAR CANE | | NEGATIVE | PASS |
| KIT INSTRUCTIONS | | REV NO. 23050 | |

REVIEWED BY:  DATE: 2/10/2025

QC APPROVAL:  DATE: 2/11/2025